

331 East Evelyn Avenue  
Mountain View, CA 94041, USA  
Main 650-962-4000  
Fax 650-962-5200  
www.conceptus.com  
www.essuremd.com  
www.essure.com

August 2009

Dear Physician:

Conceptus, Inc. would like to provide you with the following information regarding endometrial ablation procedures in patients with Essure micro-inserts in place. This labeling statement was approved by Conceptus' Notified Body, NSAI, for the purpose of CE mark approval on July 24, 2008.

**Bench and clinical studies demonstrated that endometrial ablation of the uterus can be safely and effectively performed with the Gynecare THERMACHOICE Uterine Balloon System, the Cytoc NovaSure Endometrial Ablation System, and the Boston Scientific Hydro ThermAblator immediately following Essure micro-insert placement. No specific studies have been conducted to evaluate Essure micro-insert expulsion or contraception rates following combined Essure and endometrial ablation procedures.**

An updated Physician's Instructions for Use will be available on our website in the coming months. Please feel free to contact me at [edward\\_yu@conceptus.com](mailto:edward_yu@conceptus.com) should you have any questions.

Sincerely,



Edward Yu  
Senior Director of Clinical Research and Regulatory Affairs

Information contained herein for distribution outside the USA only.